

VISERA ELITE II VIDEO SYSTEM CENTER
OTV-S200
All-in-One 2D Processor and Light Source



OTV-S200

All-in-One 2D Processor and Light Source

- Capable of 2D observation
- Compact system for simplified work flow

LCD Touch Panel

- Intuitive handling and setup
- Presets allow easy preparation and maintenance

LED Light Source

- Reduction in running costs thanks to the long lifetime of the LEDs
- Excellent natural color reproduction with the combination of an enhanced imaging process

Special Light Observation

- NBI and two modes of IR observation

Specifications			
Power Supply	Rated voltage	100–240 V AC; within $\pm 10\%$	
	Rated frequency	50/60 Hz; within ± 1 Hz	
	Rated input	400 VA	
Size	Dimensions (maximum)	383 (W) \times 199 (H) \times 506 (D) mm	
	Weight	19.3 kg	
Observation	Analog signal output	VBS composite and Y/C; simultaneous outputs possible.	
	Digital signal output	HD-SDI (SMPTE292M), DVI (WUXGA, 1080 pixels, or SXGA can be selected)	
	Electronic zoom	The image enlargement level can be selected. 3 modes (1.0 \times , 1.2 \times , 1.5 \times)	
	Optical-digital observation	The optical-digital observation can be performed. The endoscope compatible with the optical-digital observation is required.	
		NBI observation	This observation mode uses narrow-band light
Documentation	Remote control	The following ancillary equipment can be controlled (specified models only): · Portable memory / · Video recorder / · Video printer / · Image filing system	
	Recording format and number of recorded images in internal memory	TIFF: no compression	Approx. 120 images
		JPEG (1/5): Approx. 1/5 compression	Approx. 636 images
		JPEG (1/10): Approx. 1/10 compression	Approx. 1108 images
	These are the numbers of recorded images when both HDTV and SDTV images are recorded. These numbers vary depending on the images.		
Illumination	Examination lamp	LED	
	Cooling	Forced-air cooling	
	Providing maximum light intensity	Less than 5.64W (approximately same as Olympus light source that is mounted xenon short-arc lamp 300 W)	
	Observation mode	WLI or NBI observation IR observation (when connected to CLV-S200-IR)	
Automatic Brightness Adjustment	Automatic brightness adjustment method	LED drive current control	
	Automatic exposure	17 steps	
	Brightness mode	Auto Manual	
Classification (Medical Electrical Equipment)	Type of protection against electric shock	Class I	
	Degree of protection against electric shock of applied part	Depends on applied part. Also refer to applied part (camera head or videoscope).	
	Degree of protection against explosion	The video system center should be kept away from flammable gases	

Specifications, design, and accessories are subject to change without any notice or obligation on the part of the manufacturer.

OLYMPUS

OLYMPUS EUROPA SE & CO. KG
Postbox 10 49 08, 20034 Hamburg, Germany
Wendenstrasse 14–18, 20097 Hamburg, Germany
Phone: +49 40 23773-0, Fax: +49 40 233765
www.olympus-europa.com



E0492491EN · 1.000 · 05/17 · PR · ED